

NANOCOLOR® CSB 160

COD / DCO / DQO

436 nm


Method(e) / Método

0261


15 - 160 mg/l O₂

Test 0-26


REF 985 026




2.0 ml Probe
Sample
Enchantillon
Muestra




Sicherheitsgefäß
Safety bottle
Récipient de sécurité
Recipiente de seguridad




148 °C / 2 h




Schütteln
Shake
Agiter
Agitar



Abkühlen lassen
Allow to cool
Laisser refroidir
Dejar enfriar



Säubern
Clean
Nettoyer
Limpiar



Messung
Measurement
Mesure
Medición

Meerwasser
Sea water
Eau de mer
Agua de mar

