

NANOCOLOR® gesamt-Phosphat 50

Test 0-79


total Phosphate(s) / Fosfato total

REF 985 079


436 nm

Method(e) / Método


0791	10.0 - 50.0 mg/l P
0791	30 - 150 mg/l PO_4^{3-}
0793	23 - 110 mg/l P_2O_5




4.0 ml Probe
Sample
Enchantillon
Muestra




1 x NANOFIX R2




120 °C / 30 min
oder / or / ou / o
100 °C / 1 h




Abkühlen lassen
Allow to cool
Laisser refroidir
Dejar enfriar




1.0 ml R3




Schütteln
Shake
Agiter
Agitar




Säubern
Clean
Nettoyer
Limpiar



10'00 min



Messung
Measurement
Mesure
Medición



Meerwasser
Sea water
Eau de mer
Agua de mar

NANOCOLOR® ortho-Phosphat 50

ortho Phosphate(s) / orto Fosfato


Test 0-79A

REF 985 079


436 nm

Method(e) / Método


0794	10.0 - 50.0 mg/l $\text{PO}_4\text{-P}$
0795	30 - 150 mg/l PO_4^{3-}
0796	23 - 110 mg/l P_2O_5




1
4.0 ml Probe
Sample
Enchantillon
Muestra




2
1.0 ml R3




3
Schütteln
Shake
Agiter
Agitar



4
10'00 min



5
Messung
Measurement
Mesure
Medición



Meerwasser
Sea water
Eau de mer
Agua de mar