

NANOCOLOR® gesamt-Phosphat 5

total Phosphate(s) / Fosfato total


Test 0-81

REF 985 081


690 nm

Method(e) / Método


0811 0.20 - 5.00 mg/l P
0812 0.5 - 15.0 mg/l PO₄³⁻




1
1.0 ml Probe
Sample
Enchantillon
Muestra




2
1 x NANOFIX R2




3
120 °C / 30 min
oder / or / ou / o
100 °C / 1 h




4
Abkühlen lassen
Allow to cool
Laisser refroidir
Dejar enfriar




5
1 x NANOFIX R3




6
200 µl R4




7
Schütteln
Shake
Agiter
Agitar



8
Säubern
Clean
Nettoyer
Limpiar




9
10'00 min



10
Messung
Measurement
Mesure
Medición

Meerwasser
Sea water
Eau de mer
Agua de mar



NANOCOLOR® ortho-Phosphat 5

ortho Phosphate(s) / orto Fosfato


Test 0-81A

REF 985 081


690 nm

Method(e) / Método


0815 0.20 - 5.00 mg/l $\text{PO}_4\text{-P}$
0816 0.5 - 15.0 mg/l $\text{PO}_4\text{-P}$




1
1.0 ml Probe
Sample
Enchantillon
Muestra




2
1 x NANOFIX R3




3
200 µl R4




4
Schütteln
Shake
Agiter
Agitar



5
Säubern
Clean
Nettoyer
Limpiar



6
10'00 min



Messung
Measurement
Mesure
Medición

Meerwasser
Sea water
Eau de mer
Agua de mar

