

# visicolor® ECO Chlordioxid

Chlorine dioxide / Dioxyde de chlore / Cloro dióxido

**Test 5-21**


REF 931 221

540 nm


Method(e) / Método

5211


0.20–3.80 mg/L ClO<sub>2</sub>




Probe / sample  
échantillon / muestra




**1**  
5 mL Probe  
Sample  
Echantillon  
Muestra




**2**  
2  $\Delta$  ClO<sub>2</sub>-1




**3**  
Schütteln  
Shake  
Agiter  
Agitar




**4**  
2'00 min




**5**  
3  $\Delta$  ClO<sub>2</sub>-2




**6**  
3  $\Delta$  ClO<sub>2</sub>-3



**7**  
Schütteln  
Shake  
Agiter  
Agitar



**8**  
Säubern  
Clean  
Nettoyer  
Limpiar



Messung  
Measurement  
Mesure  
Medición



Meerwasser  
Sea water  
Eau de mer  
Agua de mar