

# Chlor/Ozon 2

## Chlorine/Chlore/Cloro//Ozone/Ozono

**Test 0-17**

REF 985 017



Rundküvettentest / Tube test

Test en cuves rondes / Prueba en cubeta cilíndrica

0.05–2.50 mg/L Cl<sub>2</sub>

0.09–4.80 mg/L ClO<sub>2</sub>

Freies Chlor / Free Chlorine / Chlore libre / Cloro libre:


Chlordioxid / Chlorine dioxide / Dioxyde de chlore / Cloro dióxido:

540 nm


Method(e) / Método

0171


0174




**1**  
4.0 mL Probe  
Sample  
Echantillon  
Muestra




**2**  
Kräftig schütteln  
Shake well  
Bien agiter  
Agitar intensamente




**3**  
Säubern  
Clean  
Nettoyer  
Limpiar



**4**  
1'00 min



**5**  
Lösung A  
Solution A  
Solution A  
Solución A



**6**  
Messung  
Measurement  
Mesure  
Medición

# Chlor / Ozon 2

Chlorine / Chlore / Cloro // Ozone / Ozono

**Test 0-17**

REF 985 017




0.05–2.50 mg/L Cl<sub>2</sub>


**Gesamtchlor / Total chlorine / Chlore total / Cloro total:**

Method(e) / Método


**0172**




Lösung A  
Solution A  
Solution A  
Solución A




**3 x R2**




Schütteln  
Shake  
Agiter  
Agitar



Säubern  
Clean  
Nettoyer  
Limpiar



**3'00 min**



Messung  
Measurement  
Mesure  
Medición

# Chlor/Ozon 2

Chlorine/Chlore/Cloro // Ozone/Ozono

**Test 0-17**

REF 985 017




0.05–2.00 mg/L O<sub>3</sub>


**Ozon/Ozone/Ozono:**

Method(e) / Método


**0173**




**1**  
3 x 10 R2




**2**  
5 mL Probe  
Sample  
Echantillon  
Muestra




**3**  
Schütteln  
Shake  
Agiter  
Agitar



**4**  
Säubern  
Clean  
Nettoyer  
Limpiar



**5**  
2'00 min



**6**  
Messung  
Measurement  
Mesure  
Medición