

Chlor / Ozon 2

Chlorine / Chlore / Cloro // Ozone / Ozono

Test 0-17

REF 985 017




0.05–2.50 mg/L Cl₂


Gesamtchlor / Total chlorine / Chlore total / Cloro total:

Method(e) / Método


0172




Lösung A
Solution A
Solution A
Solución A




3 x 2 R2




Schütteln
Shake
Agiter
Agitar



Säubern
Clean
Nettoyer
Limpiar



3'00 min



Messung
Measurement
Mesure
Medición