




**Rundküvettentest/ Tube test**

**Test en cuves rondes/ Prueba en cubeta cilíndrica**


50–300 mg/L

436 nm


Method(e) / Método  
**0331**




**2.0 mL** Probe  
Sample  
Echantillon  
Muestra




Sicherheitsgefäß  
Safety bottle  
Récipient de sécurité  
Recipiente de seguridad




**148 °C/2 h**




Schütteln  
Shake  
Agiter  
Agitar



Abkühlen lassen  
Allow to cool  
Laisser refroidir  
Dejar enfriar



Säubern  
Clean  
Nettoyer  
Limpiar



Messung  
Measurement  
Mesure  
Medición