

# Härte 20

## Hardness / Dureté / Durezza

**Test 0-43**

REF 985 043



Rundküvettentest / Tube test

Test en cuves rondes / Prueba en cubeta cilíndrica

1.0–20.0 °d / 0.2–3.6 mmol/L

1.0–25.0 °e / 2.0–36.0 °f


Gesamthärte / Total hardness / Dureté total / Durezza total:

540 nm


Method(e) / Método

0431/2


0433/4




1 x **NANOFIX R2**




Kräftig schütteln  
Shake well  
Bien agiter  
Agitar intensamente




2'00 min




200 µL Probe  
Sample  
Echantillon  
Muestra




Schütteln  
Shake  
Agiter  
Agitar




Säubern  
Clean  
Nettoyer  
Limpiar



1'00 min



Lösung A  
Solution A  
Solution A  
Solución A



Messung  
Measurement  
Mesure  
Medición

# Härte 20

## Hardness / Dureté / Dureza

**Test 0-43**

REF 985 043




10–100 mg/L Ca<sup>2+</sup>


Calcium / Calcio:

Method(e) / Método


**0435**




Lösung A = Null  
Solution A = Zero  
Solution A = Zéro  
Solución A = Cero




**200 µL R3**




Schütteln  
Shake  
Agiter  
Agitar



Säubern  
Clean  
Nettoyer  
Limpiar



**1'00 min**



Messung  
Measurement  
Mesure  
Medición

# Härte 20

## Hardness / Dureté / Dureza

**Test 0-43**

REF 985 043




5–50 mg/L Mg<sup>2+</sup>


**Magnesium / Magnesio:**

Method(e) / Método


**0436**




Lösung A  
Solution A  
Solution A  
Solución A




200 µL R3




Schütteln  
Shake  
Agiter  
Agitar



Säubern  
Clean  
Nettoyer  
Limpiar



1'00 min



Messung  
Measurement  
Mesure  
Medición