


Härte 20

Hardness / Dureté / Dureza

Test 0-43

REF 985 043




10–100 mg/L Ca²⁺


Calcium / Calcio:

Method(e) / Método


0435




Lösung A = Null
Solution A = Zero
Solution A = Zéro
Solución A = Cero




200 µL R3




Schütteln
Shake
Agiter
Agitar



Säubern
Clean
Nettoyer
Limpiar



1'00 min



Messung
Measurement
Mesure
Medición