

gesamt-Stickstoff TN_b 220

total Nitrogen /Azote total /Nitrogeno total

Test 0-88
REF 985 088



Rundküvettentest / Tube test


365 nm

Test en cuves rondes / Prueba en cubeta cilíndrica


Method(e) / Método

5–220 mg/L N


0881




1
Aufschlussküvette
Decomposition tube
Cuve ronde de minéralisation
Tubo de descomposición




2
500 µL Probe
Sample
Echantillon
Muestra




3
1 x orange/naranja
Aufschlussreagenz
Decomposition reagent
Réactif de minéralisation
Reactivo de descomposición




4
Schütteln
Shake
Agiter
Agitar




5
120 °C/30 min
oder/or/ou/o
100 °C/1 h




6
Abkühlen lassen
Allow to cool
Laisser refroidir
Dejar enfriar



7
1 x schwarz/black/noire/negro
Kompensationsreagenz
Compensation reagent
Réactif de compensation
Reactivo de compensación



8
Schütteln
Shake
Agiter
Agitar



9
Aufschlusslösung **A**
Decomposed solution **A**
Solution de minéralisation **A**
Solución de descomposición **A**