

gesamt-Stickstoff TN_b 60

total Nitrogen /Azote total /Nitrogeno total

Test 0-92

REF 985 092



Rundküvettentest / Tube test


Test en cuves rondes / Prueba en cubeta cilíndrica

365 nm


Method(e) / Método

0921


3–60 mg/L N




1
Aufschlussküvette
Decomposition tube
Cuve ronde de minéralisation
Tubo de descomposición




2
1.0 mL Probe
Sample
Echantillon
Muestra




3
1 x orange/naranja
Aufschlussreagenz
Decomposition reagent
Réactif de minéralisation
Reactivo de descomposición




4
Schütteln
Shake
Agiter
Agitar




5
120 °C/30 min
oder/or/ou/o
100 °C/1 h




6
Abkühlen lassen
Allow to cool
Laisser refroidir
Dejar enfriar



7
1 x schwarz/black/noire/negro
Kompensationsreagenz
Compensation reagent
Réactif de compensation
Reactivo de compensación



8
Schütteln
Shake
Agiter
Agitar



9
Aufschlusslösung **A**
Decomposed solution **A**
Solution de minéralisation **A**
Solución de descomposición **A**