


NANOCOLOR® CSB 1500 Hg-frei

COD 1500 Hg-free / DCO 1500 sans Hg / DQO 1500 sin Hg



Test 0-29
REF 963029

Methode / Method / Méthode / Método
0291 100–1500 mg/L O₂

595 / 605 / 620 nm

Min	20–80 %		Max
100	380	1220	1500 mg/L O ₂

