


NANOCOLOR® Chlor / Ozon 2

Chlorine / Ozone / Chlore / Ozon / Cloro / Ozono








Test 0-17
REF 985017

Methode / Method / Méthode / Método
0171 0.05–2.50 mg/L Cl₂
freies Chlor / free Chlorine
Chlore libre / Cloro libre


530 / 540 nm

Min	20–80 %		Max
0.05	0.54	2.01	2.50 mg/L Cl ₂



NANOCOLOR[®] Chlor / Ozon 2

Chlorine / Ozone / Chlore / Ozon / Cloro / Ozono








Test 0-17
REF 985017

Methode / Method / Méthode / Método
0172 0.05–2.50 mg/L Cl₂
Gesamtchlor / total Chlorine
Chlore total / Cloro total


530 / 540 nm

Min	20–80 %		Max
0.05	0.54	2.01	2.50 mg/L Cl ₂



NANOCOLOR® Chlor / Ozon 2

Chlorine / Ozone / Chlore / Ozone / Cloro / Ozono









Test 0-17
REF 985017

Methode / Method / Méthode / Método
0173 0.05–2.00 mg/L O₃


540 nm

Min	20–80 %		Max
0.05	0.44	1.61	2.00 mg/L O ₃



NANOCOLOR® Chlor / Ozon 2

Chlorine / Ozone / Chlore / Ozon / Cloro / Ozono



Test 0-17
REF 985017

Methode / Method / Méthode / Método
0174 0.09–4.80 mg/L ClO₂

540 nm

Min	20–80 %		Max
0.09	1.03	3.86	4.80 mg/L ClO ₂

