


NANOCOLOR[®] CSB 15000

COD/DCO/DQO



Test 0-28
REF 985028

Methode / Method / Méthode / Método
0281 1.0–15.0 g/L O₂

595 / 605 / 620 nm

Min	20–80 %		Max
1.0	3.8	12.2	15.0 g/L O ₂

