


NANOCOLOR[®] gesamt-Phosphat 1

total Phosphate / Phosphate total / Fosfato total













Test 0-76
REF 985076

Methode / Method / Méthode / Método

0761 0.05–1.50 mg/L P
0762 0.2–5.0 mg/L PO₄³⁻


690 nm

Min	20–80 %		Max
0.05	0.34	1.21	1.50 mg/L P
0.2	1.2	4.0	5.0 mg/L PO ₄ ³⁻



NANOCOLOR® ortho-Phosphat 1

ortho-Phosphate / ortho Phosphates / orto Fosfato



Test 0-76
REF 985076

Methode / Method / Méthode / Método

0765 0.05–1.50 mg/L PO₄-P

0766 0.2–5.0 mg/L PO₄³⁻

690 nm

Min	20–80 %		Max
0.05	0.34	1.21	1.50 mg/L PO ₄ -P
0.2	1.2	4.0	5.0 mg/L PO ₄ ³⁻

