




# Silica HE / Silice HE / Ácido sílicico HE




Cat. No.	Cuvette	Measurement range	Method	Wavelength
918 48	50 mm	0.002 - 0.100 mg/l Si 0.005 - 0.200 mg/l SiO <sub>2</sub>	(1) 483 (1) 484	800 nm




**1**  
20 ml sample  
échantillon  
muestra




**2**  
1 ml R1




**3**  
mix  
mélanger  
mezclar



**4**  
3'00 min




**5**  
1 ml R2




**6**  
mix  
mélanger  
mezclar




**7**  
1'00 min




**8**  
1 ml R3




**9**  
fill up with dist. water  
emplir avec l'eau dist.  
llenar con agua dest.




**10**  
mix  
mélanger  
mezclar




**11**  
15'00 min



**12**  
rinse and fill  
rincer et emplir  
lavar y llenar



**13**  
measurement  
mesure  
medición



**13**  
blank value  
zéro  
cero



**13**  
blank measurement  
zéro mesure  
cero medición