




# Copper / Cuivre / Cobre




Cat. No.	Cuvette	Measurement range	Method	Wavelength
918 53	10 mm	0.1 - 10.0 mg/l Cu <sup>2+</sup>	(1) 531	585 nm
	50 mm	0.01 - 2.00 mg/l Cu <sup>2+</sup>		




**1**  
 20 ml sample  
 échantillon  
 muestra




**2**  
 2.0 ml R1




**3**  
 mix  
 mélanger  
 mezclar




**4**  
 pH control  
 pH contrôle  
 pH control




**4a**  
 adjust pH with R1  
 ajuster pH avec R1  
 ajustar pH con R1




**5**  
 2.0 ml R2




**6**  
 mix  
 mélanger  
 mezclar




**7**  
 fill up with dist. water  
 emplir avec l'eau dist.  
 llenar con agua dest.



**8**  
 mix  
 mélanger  
 mezclar




**9**  
 15'00 min



**10**  
 rinse and fill  
 rincer et emplir  
 lavar y llenar



**11**  
 measurement  
 mesure  
 medición



blank value  
 zéro  
 cero



blank measurement  
 zéro mesure  
 cero medición