




Nitrate Z / Nitrato Z




Cat. No.	Cuvette	Measurement range	Method	Wavelength
918 63	10 mm	0.1 - 1.0 mg/l NO ₃ - N 0.5 - 5.0 mg/l NO ₃ ⁻	(1) 631 (1) 632	520 nm
	50 mm	0.02 - 0.45 mg/l NO ₃ - N 0.1 - 2.0 mg/l NO ₃ ⁻	(1) 631 (1) 632	




1
20 ml sample
échantillon
muestra




2
1.0 ml R1




3
mix
mélanger
mezclar




4
1 drop R2




5
15 - 30 s shake
agiter
agitar




6
fill up with dist. water
emplir avec l'eau dist.
llenar con agua dest.




7
mix
mélanger
mezclar




8
10'00 min




9
rinse and fill
rincer et emplir
lavar y llenar



10 **M**
measurement
mesure
medición



10 **N**
blank value
zéro
cero



10 **N**
blank measurement
zéro mesure
cero medición