




ortho-Phosphate / ortho-Fosfato




Cat. No.	Cuvette	Measurement range	Method	Wavelength
918 77	10 mm	0.2 - 6.5 mg/l PO ₄ - P 0.5 - 20.0 mg/l PO ₄ ³⁻	(1) 771 (1) 772	690 nm
	50 mm	0.04 - 1.70 mg/l PO ₄ - P 0.1 - 5.0 mg/l PO ₄ ³⁻	(1) 771 (1) 772	




1
20 ml sample
échantillon
muestra




2
1.0 ml R1




3
mix
mélanger
mezclar




4
1.0 ml R2



5
mix
mélanger
mezclar




6
fill up with dist. water
emplir avec l'eau dist.
llenar con agua dest.




7
mix
mélanger
mezclar




8
10'00 min




9
rinse and fill
rincer et emplir
lavar y llenar



10 **M**
measurement
mesure
medición



10 **N**
blank value
zéro
cero



10 **N**
blank measurement
zéro mesure
cero medición